



Urgent Field Safety Notice

MY-FSN-RDS-Pathology Lab-2022-006

RDS / Pathology Lab / VENTANA PD-L1 (SP142) Assay

Jasmine Michael

Version 1

22 November 2022

Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots

Product Name	VENTANA PD-L1 (SP142) Assay (Class III, US IVD) VENTANA PD-L1 (SP142) Assay (CE IVD Predictive) VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)
GMMI / Part No	VENTANA PD-L1 (SP142) Assay (Class III, US IVD)
Device Identifier	GMMI: 07709374001, UDI: 04015630967155 VENTANA PD-L1 (SP142) Assay (CE IVD Predictive) GMMI: 08008540001 UDI: 07613336106709 VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical) GMMI: 07011571001 UDI: 04015630976591
Production Identifier (Lot No. /Serial No.)	Please refer to Table A and Table B.
SW Version	Not applicable
Type of Action	Field Safety Corrective Action (FSCA)

Dear Valued Customer,

Description of Situation

VENTANA Medical Systems, Inc. (Roche) observed unacceptable, light staining with some VENTANA PD-L1 (SP142) on-market lots, during internal comparison studies. Light staining affects the borderline of positive versus negative test results. It can cause a reduction in both staining intensity and the percent positivity, which could potentially lead to a false-negative PD-L1 result. In addition, where the assay is used as a companion diagnostic, a patient may not be identified as eligible for treatment with the associated therapy TECENTRIQ® (atezolizumab).

Considering the hazard involved, there is a remote probability that use of the affected lots is likely to lead to adverse events in populations at greatest risk such as patients undergoing testing with Ventana PD-L1 SP142



Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots

assay for treatment with TECENTRIQ® (atezolizumab). Adverse health consequences are otherwise not likely.

VENTANA Medical Systems, Inc. (Roche) is requesting that customers identify, discontinue the use of and discard any remaining inventory of the lots listed below in Table A. Additionally; the lots listed in Table B are being short dated to ensure optimum product performance.

Please refer to Table A and Table B below for specific information and additional details.

Actions taken by Roche Diagnostics

The on-going investigation has determined the root cause is related to variability in the selection of antibody concentration in raw materials, affecting specific VENTANA PD-L1 (SP142) Assay lots made with the impacted raw materials. A CAPA investigation is still ongoing to further identify and address the root cause.

Actions to be taken by the customer/user

Customers should immediately discontinue the use of and discard any remaining inventory of the VENTANA PD-L1 (SP142) Assay specific lots listed in Table A below.

Additionally, other specific lots are being short dated to ensure optimum product performance in the field; refer to Table B below. These short dated lots can be used until the updated Date of Expiration.

Furthermore, all negative results generated using specific, affected lots are to be reviewed.

Customers should follow their standard laboratory operating procedures to investigate any suspected false negative results generated with the affected lots.

Please find the list of VENTANA PD-L1 (SP142) Assay lots and related information in Table A and Table B.

Table A

The usage of the following affected VENTANA PD-L1 (SP142) Assay lots must be immediately discontinued and the dispensers must be discarded.

Description	Material Number	Affected Lots	Date of Expiration (MM-DD-YYYY)	Reached Expiration	Negative results review needed
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	F30596	01/01/2022	Yes	Yes
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	G31438	12/09/2022	No	Yes
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	H03958	02/23/2023	No	Yes
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	F31427	01/02/2022	Yes	Yes
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	G06715	03/29/2022	Yes	Yes
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	G16831	06/17/2022	Yes	Yes
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	G24564	09/13/2022	Yes	Yes
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	H00001	01/10/2023	No	Yes
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	H07164	03/22/2023	No	Yes

Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots

VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)	07011571001	H08373	04/12/2023	No	Yes
---	-------------	--------	------------	----	-----

Table B

The following affected VENTANA PD-L1 (SP142) Assay lots are being short dated. Please refer to the Current and Updated Dates of Expiration and the column indicating whether previously generated negative results should be reviewed. If the "Updated Dates of Expiration" is in the past, these lots should no longer be distributed nor should they be used by customers.

Description	Material Number	Affected Lots	Current Date of Expiration (MM-DD-YYYY)	Updated Date of Expiration (MM-DD-YYYY)	Negative results review needed
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	G27114	11/01/2022	03/01/2022	Yes - Only results generated after 03/01/2022
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	H13049	05/09/2023	09/09/2022	Yes - Only results generated after 09/09/2022
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	H35110	01/06/2024	05/06/2023	No
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	J25361	09/29/2024	01/29/2024	No
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	J12769	05/03/2024	09/03/2023	No
VENTANA PD-L1 (SP142) Assay (Class III, US IVD)	07709374001	J18775	07/25/2024	11/25/2023	No
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	H15132	05/27/2023	09/27/2022	Yes - Only results generated after 09/27/2022
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	H18578	07/18/2023	11/18/2022	No
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	H23050	08/29/2023	12/29/2022	No
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	J00572	01/06/2024	05/06/2023	No
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	J01568	02/03/2024	06/03/2023	No
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	J14244	05/30/2024	09/30/2023	No
VENTANA PD-L1 (SP142) Assay (CE IVD Predictive)	08008540001	J20989	08/07/2024	12/07/2023	No
VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)	07011571001	H17865	07/08/2023	11/08/2022	Yes - Only results generated after 11/08/2022
VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)	07011571001	H30521	11/04/2023	03/04/2023	No
VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)	07011571001	H35111	01/09/2024	05/09/2023	No
VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)	07011571001	J11932	05/01/2024	09/01/2023	No
VENTANA PD-L1 (SP142) IHC Assay (CE IVD Analytical)	07011571001	J20132	08/07/2024	12/07/2023	No

Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots

Communication of this Field Safety Notice (if appropriate)

This notice must be passed on to all those who need to be aware within your organization or to any organization/individual where the potentially affected devices have been distributed/supplied. (If appropriate).

Please transfer this notice to other organizations/individuals on which this action has an impact. (If appropriate).

Please maintain awareness of this notice and resulting action for an appropriate period to ensure the effectiveness of the corrective action. (If appropriate).

Please complete and return the **Customer Acknowledgement Receipt** within 2 business days to acknowledge your reading and understanding of this notice.

We apologize for any inconvenience this may cause and hope for your understanding and your support.

Contact

If you have any questions, please do not hesitate to contact your sales representative or our Customer Call Centre: **1800 88 8881**

Thank you.

Best regards,

ROCHE DIAGNOSTICS (M) SDN. BHD.



Jasmine Michael

MS Marketing Partner

Email: jasmine.michael@roche.com

Tel. +603-7967 2388



Ong Yean Ting

Head, Regulatory Affairs and Quality

Email: yeon_ting.ong@roche.com

Tel. +603-7967 2390



Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots

Customer Acknowledgement Receipt

Roche Diagnostics Malaysia
Fax Back: +603 7967 2399
Email: malaysia.diagnostics@roche.com

Attention: Pathology Laboratory & Customer Support Team

Re: MY-FSN-RDS-Pathology Lab-2022-006 Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots

I acknowledge I have received notification on MY-FSN-RDS-Pathology Lab-2022-006 Unacceptable Light Staining & review of negative results - VENTANA PD-L1 (SP142) Assay Specific Lots and will proceed to take the action suggested i.e. take note of the corrected expiration date and discard all lots that have gone past the corrected expiration date.

No. of kits disposed: _____

Thank you.

Name:

Organization:

Stamp:

Date: