

**MDA/GL/06**

**June 2025**

**Forth Edition**

# **GUIDELINE ON HOW TO APPLY FOR PRODUCT CLASSIFICATION APPLICATION**



**Medical Device Authority**  
MINISTRY OF HEALTH MALAYSIA

| <b>Contents</b>                 | <b>Page</b> |
|---------------------------------|-------------|
| Preface .....                   | iii         |
| 0 Introduction .....            | 1           |
| 1 Scope and application .....   | 2           |
| 2 Terms and definitions .....   | 2           |
| 3 Classification Criteria ..... | 3           |
| 4 Application Procedure .....   | 3           |

## Preface

This Guideline Document was prepared by the Medical Device Authority (MDA) to help the industry and healthcare professionals in their quest to comply with the Medical Device Act (Act 737) and the regulations under it.

This Guideline Document shall be read in conjunction with the current laws and regulations used in Malaysia, which include but not limited to the following-

- a) Medical Device Act 2012 (Act 737); and
- b) Medical Device Regulations 2012;

In this Guideline Document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission; and
- “can” indicates a possibility or a capability.

When a requirement is required to be “documented”, it is also required to be established, implemented and maintained.

Irrespective of the requirements of this Guideline Document, MDA has the right to request for information or material, or define conditions not specifically described in this document that is deemed necessary for the purpose of regulatory control.

MDA has put much effort to ensure the accuracy and completeness of this guideline document. In the event of any contradiction between the contents of this document and any written law, the latter should take precedence.

MDA reserves the right to amend any part of the guideline document from time to time.

## CONTACT INFORMATION

For further information, please contact:

**UNIT PENKELASAN DAN KOMBINASI PRODUK  
CAWANGAN PENDAFTARAN  
BAHAGIAN KAWALAN PRA-PASARAN  
MEDICAL DEVICE AUTHORITY**

Ministry of Health Malaysia  
Level 6, Prima 9, Prima Avenue II  
Block 3547, Persiaran APEC  
63000 Cyberjaya, Selangor  
MALAYSIA  
T: (03) 8230 0300  
Website: [portal.mda.gov.my](http://portal.mda.gov.my)  
Email: [classification@mda.gov.my](mailto:classification@mda.gov.my)

# **GUIDELINE ON HOW ~~TO SUBMIT AN APPLICATION FOR PRODUCT TO~~ APPLY FOR PRODUCT CLASSIFICATION APPLICATION**

## **0 Introduction**

Section 3(1) Act 737, a medical device shall be classified by an establishment based on the level of risk it poses, its intended use and the vulnerability of the human body in accordance with the prescribed manner.

Section 3(2) in the event of any dispute between an establishment and a conformity assessment body over a classification of a medical device, the matter shall be referred to the Authority, in the manner and within such period as may be specified by the Authority, for its decision.

Circular letter of the Medical Device Authority No. 5 Year 2016, the policy on implementation and enforcement under the Medical Device Act 2012 (Act 737) has been released upon the imposition of charges or fees for product classification.

Many manufacturers have difficulty in interpreting whether or not their product would be considered a medical device within the terms of the Malaysia Medical Device Regulations 2012 (Act 737). This guideline document has been developed to aid with some of the more common areas of confusion.

It is often assumed that because a product is considered a medical device in some countries, for example in the USA, EU, Canada or in Japan, that it will also be a medical device in the Malaysia. This is not the case and manufacturers should always refer to the Malaysia definitions of a medical device when making any borderline determinations. Any such decision will be based on the stated intended purpose of the product and its mode of action. Manufacturers should also consult the available published guidance in order to determine whether or not their product is considered a medical device in the Malaysia.

In general, medical devices must have a 'medical purpose' which is determined by the definition of a medical device. They must also act primarily in a way that is not metabolic, immunological or pharmacological. Should they function in any way that is metabolic, immunological or pharmacological, in conjunction with having a medical purpose, they are likely to come within the remit of the regulations covering medicinal products instead. Further information on the borderline products is available – refer to **Guidance Document of Harmonised Borderline Products in ASEAN**

## 1 Scope and objective

This guideline provides guidance, reference and clarification on how to apply for Product Classification that are regulated under the Medical Device Act (Act 737). This document is applicable to establishments, healthcare facilities, and public dealing medical device and non-medical device products.

Product Classification **main objective** is **to determine whether a product is classified as a medical device product or not** under the Medical Device Act 737.

In 2025, **Product Classification** has **two main processes** which consist of **New Application of Product Classification** and **Renewal Application of Product Classification**.

## 2 Terms and definitions

For the purposes of this document, the terms and definitions in Act 737, the regulations under it, and the following apply.

### 2.1 Medical Device

Any instrument, apparatus, implement, machine, appliance, implant, in vitro reagent or calibrator, software, material or other similar or related article intended by the manufacturer to be used, alone or in combination, for human beings for the purpose of

-

- i. diagnosis, prevention, monitoring, treatment or alleviation of disease;
- ii. diagnosis, monitoring, treatment, alleviation of or compensation for an injury;
- iii. investigation, replacement or modification, or support of the anatomy or of a physiological process;
- iv. support or sustaining life;
- v. control of conception;
- vi. disinfection of medical device; or
- vii. providing information for medical or diagnostic purpose by means of in vitro examination of specimens derived from the human body,

which does not achieve its primary intended action in or on the human body by pharmacological, immunological or metabolic means, but that may be assisted in its intended function by such means.

### 2.2 Manufacturer

A person who is responsible for –

- i. The design, the production, fabrication, assembly, processing, packaging and labelling of a medical device whether or not it is the person, or a subcontractor acting on the person's behalf, who carries out these operations; and
- ii. Assigning to the finished medical device under his own name, its intended purpose and ensuring the finished products meets the regulatory requirement; or

Any other person who –

- i. Assembles, packages, processes, fully refurbishes, reprocesses or labels one or more ready-made medical devices; and
- ii. Assigning to the ready-made medical device under his own name, its intended purpose and ensuring the finished product meets the regulatory requirement,

But shall not include the following persons:

- i. Any person who assembles or adapts medical devices in the market that are intended for individual patients; and
- ii. Any person who assembles, packages or adapts medical devices in relation to which the assembling, packaging or adaptation does not change the purpose intended for the medical devices.

### **3 Classification Criteria**


- 3.1** The description and primary intended purpose of the product.
- 3.2** The primary mode of action/ the principal mechanism of action by which the claimed effect or purpose of the product is achieved by:
- 3.3** Medical device is based on function by physical means e.g.: mechanical action, creation of a physical barrier or replacement or support of organ or body function.
- 3.4** Drug is based on pharmacological, immunological or metabolic action in/on the body. (Refer [Definition of Medical Device: Guidance Document MDA/GD/0006](#))
- 3.5** Active ingredient, indication and pharmaceutical dosage form (those are the main criteria for classification of the drugs), kindly refer to Annex B: [Medical Device – Drug – Cosmetic Interphase \(MDDCI\)](#) products Table I.
- 3.6** Classification of the product/combination product or similar product/similar combination product in the reference countries. The reference countries are US, EU, UK, Canada, Australia and Japan.
- 3.7** The primary mode of action/the principal mechanism of action may be deduced from the scientific data and the manufacturer’s labelling and claims. The claims made for a product, in accordance with its mode of action may represent an important factor for its qualification as a medical device.

## 4 Application Procedure (New Application)

- 4.1 New application is to be submitted by filling in the Product Classification Application Form through the Product Classification Section in our portal website, [Product Classification Application](#)
- 4.2 Applicant is required to submit completed copies of the following documentation:
- i. **Product Information** on intended use, mode of action
  - ii. **Product Label** (indicating product name and manufacturer);
  - iii. **Product leaflet / brochure / catalogue** (that contain description, intended use);
  - iv. **Other information**, eg: User manual, Instruction for use, Packaging Insert, Declaration of Conformity, Quality Management System (QMS) Certificate, Pre-market Approval;
  - v. **Manufacturing Process (For Human Tissue Based Products)**.
- 4.3 Circular Letter of the Medical Device Authority No. 5 Year 2016: The Medical Device Authority Meeting No. 3/2016 has decided to set the policy for imposition of charges or fees for product classification. The product classification will be charged **RM300.00 per application** effective on 1st December 2016.
- 4.4 All fees shall be made through **BayarNow system** (<http://bayarnow.mda.gov.my>) **CASH WILL NOT BE accepted**. We will not be responsible for the cash sent or brought to MDA.
- 4.5 Time frame for processing application form is within **30 working days (not including weekends and public holidays)** based on the complexity of the device after the date of payment cleared. However, the time frame may take longer if the product needs to be further discussed.
- 4.6 Please be informed that the invoice issued in relation to this application is valid for a period of one (1) month from the date of issuance. If payment is not received within this timeframe, we will follow up via email to confirm whether you wish to proceed.

Should we receive confirmation to withdraw, or if no response is received within the stipulated follow-up period, the application **may be withdrawn accordingly**.

4.7 New Product Classification application process steps are as per process flow below:




## 4.7 The table below provides explanation on the above flowchart.

| Steps | Explanatory Notes  |
|-------|--|
| 1.    | The applicant shall submit a complete Application Form for Product Classification Application together with required supporting documents.<br><i>Note: The Application Form must be <b>submitted via Google Forms</b> only</i>   |
| 2.    | Pre-Market Control Division, MDA will receive the application and send an acknowledgement email to the applicant.  |
| 3.    | Pre-Market Control Division, MDA will issue a payment advice to the Finance Unit (MDA) and the Finance Unit will issue an invoice for the payment fee to the applicant via email.<br><i>Note: It will take 3 to 5 working days. <b>The payment advice and invoice will only be sent to the applicant's email address stated in the application form.</b></i> |
| 4.    | The application will be evaluated by the Pre-Market Control Division, MDA once the payment has been received and cleared.<br><i>Note: The <b>process takes within 30 working days</b> based on the complexity of the device <b>from the date of payment cleared upon complete application.</b></i>   |
| 5.    | The <b>application will be returned to the applicant through email</b> during evaluation process if the application is incomplete as the supporting documents is not satisfactory and insufficient of information. <b>The timeline for the process will be reset.</b>  |
| 6.    | The applicant will receive the Product Classification letter once the evaluation and verification process has been completed.  |
| 7.    | The Product Classification letter will be issued by the Authority via email<br><i>Note: The Product Classification letter will only be sent to <b>the applicant's email address stated in the application form.</b></i>  |

## 5 Application Procedure (Renewal Application)

- 5.1 Renewal application is to be submitted by filling in the Product Classification Application Form through the Product Classification Section in our portal website, [Product Classification Application](#)
- 5.2 Applicant is required to submit completed copies of the following documentation:
- vi. **Product Information** on intended use, mode of action
  - vii. **Product Label** (indicating product name and manufacturer);
  - viii. **Product leaflet / brochure / catalogue** (that contain description, intended use);
  - ix. **Other information**, eg: User manual, Instruction for use, Packaging Insert, Declaration of Conformity, Quality Management System (QMS) Certificate, Pre-market Approval;
  - x. **Manufacturing Process (For Human Tissue Based Products).**
- 5.3 During the transition period (2025 – 2027), the renewal application shall be process without any fee till new process will be amend in 2027.
- 5.4 Time frame for processing application form is within **14 working days (not including weekends and public holidays)** based on the complexity of the device after the date of payment cleared. However, the time frame may take longer if the product needs to be further discussed. The timeline will be based on the acknowledgement email sent to the applicant.

5.5 Renewal of Product Classification application process steps are as per process flow below:



5.6 The table below provides explanation on the above flowchart.

| Steps | Explanatory Notes   |
|-------|---|
| 1.    | The applicant shall submit a complete Application Form for Product Classification Application together with required supporting documents.<br><i>Note: The Application Form must be <b>submitted via Google Forms</b> only</i>  |
| 2.    | Pre-Market Control Division, MDA will receive the application and send an acknowledgement email to the applicant.   |
| 3.    | The details of the acknowledgement email have stated the application will proceed without invoice, only contain how many working days for the process to be done.   |
| 4.    | The application will be evaluated by the Pre-Market Control Division, MDA once the payment has been received and cleared.<br><i>Note: The <b>process takes within 14 working days</b> based on the complexity of the device <b>from the date of acknowledgement email sent.</b></i> |
| 5.    | The <b>application will be returned to the applicant through email</b> during evaluation process if the application is incomplete as the supporting documents is not satisfactory and insufficient of information. <b>The timeline for the process will be reset.</b>               |
| 6.    | The applicant will receive the Product Classification letter once the evaluation and verification process has been completed.   |
| 7.    | The Product Classification letter will be issued by the Authority via email<br><i>Note: The Product Classification letter will only be sent to <b>the applicant's email address stated in the application form.</b></i>   |

# **MEDICAL DEVICE AUTHORITY**

---

## **MINISTRY OF HEALTH, MALAYSIA**

### **Contact Information:**

**MEDICAL DEVICE AUTHORITY**  
Ministry of Health Malaysia  
Level 6, Prima 9, Prima Avenue II  
Block 3547, Persiaran APEC  
63000 Cyberjaya, Selangor  
MALAYSIA  
T: (03) 8230 0300  
F: (03) 8230 0200  
Website: [portal.mda.gov.my](http://portal.mda.gov.my)

