

FIELD CORRECTIVE ACTION NOTIFICATION AND REPORT FORM

Rev 1: September 2018
FSN Ref: FSN-2021-012 FSCA Ref: FSN-2021-012

Urgent Field Safety Notice (FSN)

Thermo Scientific™ Remel™ Shigella dysenteriae polyvalent


1. Information on Affected Devices*	
1.	1. Device Type(s)* IVD
1.	2. Commercial name(s) Thermo Scientific™ Remel™ Shigella dysenteriae polyvalent
1.	3. Unique Device Identifier(s) (UDI-DI) 05056080500904
1.	4. Primary clinical purpose of device(s)* Shigella Polyvalent Agglutinating Sera are suitable for use in slide agglutination tests to identify Shigella cultures presumptively for epidemiological and diagnostic purposes. Antisera provide serological identification only; full identification of an organism must be made in conjunction with biochemical testing.
1.	5. Device Model/Catalogue/part number(s)* R30163701
1.	6. Software version N/A
1.	7. Affected serial or lot number range 2971158
1.	8. Associated devices N/A

2. Reason for Field Safety Corrective Action (FSCA)*	
2.	1. Description of the product problem* An internal technical investigation has determined that R30163701 Shigella dysenteriae polyvalent Lot. 2971158 is failing to agglutinate with Sero type 9 and therefore lead to false negative results.
2.	2. Hazard giving rise to the FSCA* False Negatives
2.	3. Probability of problem arising High
2.	4. Predicted risk to patient/users Quality Control testing using Shigella dysenteriae group 9 positive cultures may identify this failure mode before use on clinical samples. We believe the clinical risk is extremely low based on the following rationale: Patient management should not be affected by a failure to identify the correct serotype. Patients would be treated and managed in the same way long before any serotype of Shigella dysenteriae was recovered. The risks associated with incorrect identification of the serotypes are related primarily to epidemiology, specifically that a suspected outbreak might not be linked correctly. However, Serotype 9 is rarely, if ever, recovered from clinical specimens, even in area where shigellosis is common.

FIELD CORRECTIVE ACTION NOTIFICATION AND REPORT FORM

Rev 1: September 2018

FSN Ref: FSN-2021-012 FSCA Ref: FSN-2021-012

4. General Information*		
4.	1. FSN Type*	New
4.	2. For updated FSN, reference number and date of previous FSN	N/A
4.	3. For Updated FSN, key new information as follows:	N/A
4.	4. Further advice or information already expected in follow-up FSN? *	Not planned yet
4	5. If follow-up FSN expected, what is the further advice expected to relate to:	N/A
4	6. Anticipated timescale for follow-up FSN	N/A
4.	7. Manufacturer information (For contact details of local representative refer to page 1 of this FSN)	
	a. Company Name	Thermo Fisher Scientific
	b. Address	Clipper Boulevard West, Cross ways industrial estate, Dartford, Kent. DA2 6PT
	c. Website address	www.thermofisher.com
4.	8. The Competent (Regulatory) Authority of your country has been informed about this communication to customers. *	
4.	9. List of attachments/appendices:	Customer Response Form
4.	10. Name	James Filer Vice President, Quality and Regulatory, MBD
	Signature	

Transmission of this Field Safety Notice	
	<p>This notice needs to be passed on all those who need to be aware within your organisation or to any organisation where the potentially affected devices have been transferred. (As appropriate).</p> <p>Please transfer this notice to other organisations on which this action has an impact. (As appropriate).</p> <p>Please maintain awareness on this notice and resulting action for an appropriate period to ensure effectiveness of the corrective action.</p> <p>Please report all device-related incidents to the manufacturer, distributor or local representative, and the national Competent Authority if appropriate, as this provides important feedback.*</p>

FIELD CORRECTIVE ACTION NOTIFICATION AND REPORT FORM

Rev 1: July 2018

Customer Reply Form

1. Field Safety Notice (FSN) information			
FSN Reference number*	2021-012		
FSN Date*	3 November 2021		
Product/ Device name*	ThermoScientific™ Remel™ Shigella dysenteriae polyvalent		
Product Code(s)	R30163701		
Batch/Serial Number (s)	2971158		
2. Customer Details			
Account Number			
Organisation Name*			
Organisation Address*			
Department/Unit			
Shipping address if different to above			
Contact Name*			
Title or Function			
Telephone number*			
Email*			
3. Customer action undertaken on behalf of Healthcare Organisation			
<input type="checkbox"/>	I confirm receipt of the Field Safety Notice and that I read and understood its content.		
<input type="checkbox"/>	I performed all actions requested by the FSN.		
<input type="checkbox"/>	The information and required actions have been brought to the attention of all relevant users and executed.		
<input type="checkbox"/>	I have returned affected devices - enter number of devices returned and date complete or N/A	Qty:	Lot/Serial Number: Date Returned (DD/MM/YY)
		Comments:	
<input type="checkbox"/>	I have destroyed affected devices – enter number destroyed and date complete.	Qty:	Lot/Serial Number: Date Returned (DD/MM/YY)
		Qty	Credit <input type="checkbox"/> Replacement <input type="checkbox"/>
		Comments:	
<input type="checkbox"/>	No affected devices are available for return/ destruction		
<input type="checkbox"/>	Other Action (Define):		
<input type="checkbox"/>	I do not have any affected devices.		
<input type="checkbox"/>	I have a query please contact me (e.g. need for replacement of the product).		
Print Name*			
Signature*			
Date*			

FIELD CORRECTIVE ACTION NOTIFICATION AND REPORT FORM

Rev 1: July 2018

4. Return acknowledgement to sender	
Email	MBD.vigilance@thermofisher.com
Telephone Number & Fax	Tel: +44(0) 1256 841144 Fax: +44(0) 1256 479525
Postal Address	
Deadline for returning the reply form*	30 November 2021

Mandatory fields are marked with *

It is important that your organisation takes the actions detailed in the FSN and confirms that you have received the FSN.

Your organisation's reply is the evidence we need to monitor the progress of the corrective actions.